

**4620.4550 AIR QUALITY MEASURING DEVICES.**

Subpart 1. **Device requirements.** The owner or operator must demonstrate that the device and methods used to measure air quality conditions are accurate and reliable. Air quality measuring devices must be:

A. capable of measuring carbon monoxide air concentrations in a range from 0 to 100 parts per million (ppm) in increments of 1 ppm; or

B. capable of measuring nitrogen dioxide air concentrations in a range from 0 to 5 parts per million (ppm) in increments of 0.1 ppm.

Subp. 2. **Maintenance.** The owner or operator must operate, store, maintain, and calibrate the devices according to the device manufacturer's specifications. The owner or operator must also keep maintenance and calibration records.

**Statutory Authority:** *MS s 144.1222*

**History:** *37 SR 1663*

**Published Electronically:** *June 3, 2013*