

4731.4270 RECORDS; INSPECTION AND MAINTENANCE.

Subpart 1. **Inspection and maintenance records.** A licensee must maintain records specified under part 4731.4090 of equipment problems found in daily checks and quarterly inspections of radiographic exposure devices, transport and storage containers, associated equipment, source changers, and survey instruments and retain each record for three years after it is made.

Subp. 2. **Record requirements.** The records under subpart 1 must include:

- A. the date of check or inspection;
- B. the name of inspector;
- C. equipment involved;
- D. any problems found; and
- E. what repair or maintenance, if any, was done.

Statutory Authority: *MS s 144.1202; 144.1203*

History: *29 SR 755*

Published Electronically: *March 12, 2009*