

4731.4423 AUTHORIZATION FOR CALIBRATION, TRANSMISSION, AND REFERENCE USE.

A person authorized under part 4731.4403, subpart 1, for medical use of radioactive material may receive, possess, and use the following radioactive material for check, calibration, transmission, and reference use:

A. sealed sources that do not exceed 30 millicuries (1.11 GBq) each and that are manufactured and distributed by a person licensed under part 4731.3400 or equivalent requirements of the NRC or an agreement state;

B. sealed sources that do not exceed 30 millicuries (1.11 GBq) each and that are redistributed by a licensee authorized to redistribute the sealed sources manufactured and distributed by a person licensed under part 4731.3400, providing the redistributed sealed sources are in the original packaging and shielding and are accompanied by the manufacturer's approved instructions;

C. any radioactive material with a half-life not longer than 120 days in individual amounts not to exceed 15 millicuries (0.56 GBq);

D. any radioactive material with a half-life longer than 120 days in individual amounts not to exceed the smaller of 200 microcuries (7.4 MBq) or 1,000 times the quantities in part 4731.3160; and

E. technetium-99m in amounts as needed.

Statutory Authority: *MS s 144.1202; 144.1203*

History: *29 SR 755; 32 SR 831*

Published Electronically: *March 12, 2009*