

**4731.4451 SURVEYS AFTER SOURCE IMPLANT AND REMOVAL.**

A. Immediately after implanting sources in a patient or human research subject, a licensee must make a survey to locate and account for all sources that have not been implanted.

B. Immediately after removing the last temporary implant source from a patient or human research subject, a licensee must make a survey of the patient or human research subject with a radiation detection survey instrument to confirm that all sources have been removed.

C. A licensee must retain a record of the surveys required under this part according to part 4731.4511.

**Statutory Authority:** *MS s 144.1202; 144.1203*

**History:** *29 SR 755*

**Published Electronically:** *March 12, 2009*