

4731.4456 DECAY OF STRONTIUM-90 SOURCES FOR OPHTHALMIC TREATMENTS.

A. Only an authorized medical physicist shall calculate the activity of each strontium-90 source that is used to determine the treatment times for ophthalmic treatments. The decay must be based on the activity determined under part 4731.4455.

B. A licensee must maintain a record of the activity of each strontium-90 source according to part 4731.4514.

Statutory Authority: *MS s 144.1202; 144.1203*

History: *29 SR 755*

Published Electronically: *March 12, 2009*