

4731.4457 THERAPY-RELATED COMPUTER SYSTEMS.

A licensee must perform acceptance testing on the treatment planning system of therapy-related computer systems according to published protocols accepted by nationally recognized bodies. At a minimum, the acceptance testing must include, as applicable, verification of:

- A. the source-specific input parameters required by the dose calculation algorithm;
- B. the accuracy of dose, dwell time, and treatment time calculations at representative points;
- C. the accuracy of isodose plots and graphic displays; and
- D. the accuracy of the software used to determine sealed source positions from radiographic images.

Statutory Authority: *MS s 144.1202; 144.1203*

History: *29 SR 755*

Published Electronically: *March 12, 2009*