

**4731.4461 USE OF SEALED SOURCES FOR DIAGNOSIS; TRAINING.**

Except as provided under part 4731.4414, a licensee must require an authorized user of a diagnostic sealed source for use in a device authorized under part 4731.4460 to be a physician, dentist, or podiatrist who:

A. is certified by a specialty board whose certification process includes all of the requirements of item B and whose certification has been recognized by the commissioner, the NRC, or an agreement state; or

B. has:

(1) completed eight hours of classroom and laboratory training in basic radionuclide handling techniques specifically applicable to the use of the device. The training must include:

(a) radiation physics and instrumentation;

(b) radiation protection;

(c) mathematics pertaining to the use and measurement of radioactivity;

and

(d) radiation biology; and

(2) completed training in the use of the device for the uses requested.

**Statutory Authority:** *MS s 144.1202; 144.1203*

**History:** *29 SR 755; 32 SR 831*

**Published Electronically:** *March 12, 2009*