

**4731.4502 INSTRUMENT CALIBRATION RECORDS.**

Subpart 1. **Activity measurement instruments.** A licensee must maintain a record of instrument calibrations required under part 4731.4420 for three years. The record must include:

- A. the model and serial numbers of the instrument;
- B. the date of the calibration;
- C. the results of the calibration; and
- D. the name of the individual who performed the calibration.

Subp. 2. **Survey instruments.** A licensee must maintain a record of radiation survey instrument calibrations required under part 4731.4421 for three years. The record must include:

- A. the model and serial number of the instrument;
- B. the date of the calibration;
- C. the results of the calibration; and
- D. the name of the individual who performed the calibration.

**Statutory Authority:** *MS s 144.1202; 144.1203*

**History:** *29 SR 755*

**Published Electronically:** *March 12, 2009*