

4731.4504 LEAKS TEST AND INVENTORY RECORDS.

Subpart 1. **Leak tests.** A licensee must retain records of leak tests required under part 4731.4424, item B, for three years. The records must contain:

- A. the model number and serial number, if one has been assigned, of each source tested;
- B. the identity of each source radionuclide and its estimated activity;
- C. the results of the test;
- D. the date of the test; and
- E. the name of the individual who performed the test.

Subp. 2. **Inventories.** A licensee must retain records of the semiannual physical inventory of sealed sources and brachytherapy sources required under part 4731.4424, item G, for three years. The inventory records must contain:

- A. the model number and serial number, if one has been assigned, of each source;
- B. the identity of each source radionuclide and its nominal activity;
- C. the location of each source; and
- D. the name of the individual who performed the inventory.

Statutory Authority: *MS s 144.1202; 144.1203*

History: *29 SR 755*

Published Electronically: *March 12, 2009*