REVISOR

4731.4514 STRONTIUM-90 DECAY RECORDS.

A licensee must maintain a record of the activity of a strontium-90 source required under part 4731.4456 for the life of the source. The record must include:

A. the date and initial activity of the source as determined under part 4731.4455; and

B. for each decay calculation, the date and the source activity as determined under part 4731.4456.

Statutory Authority: MS s 144.1202; 144.1203

History: 29 SR 755

Published Electronically: March 12, 2009