

4731.4523 SURVEY RECORDS; THERAPEUTIC TREATMENT UNITS.

A licensee must maintain a record of radiation surveys of treatment units made according to part 4731.4476 for the duration of use of the unit. The record must include:

- A. the date of the measurements;
- B. the manufacturer's name, model number, and serial number for the treatment unit, source, and instrument used to measure radiation levels;
- C. each dose rate measured around the source while the unit is in the off position and the average of all measurements; and
- D. the signature of the individual who performed the test.

Statutory Authority: *MS s 144.1202; 144.1203*

History: *29 SR 755*

Published Electronically: *March 12, 2009*