

**4732.0630 BYPASSING A SAFETY DEVICE.**

The registrant must ensure that:

A. a safety device or interlock is not bypassed unless written approval has been obtained from the radiation safety officer or an alternate designated by the radiation safety officer:

- (1) is recorded in a permanent log; and
- (2) is for a specified period of time;

B. the bypass or safety interlock must be terminated as soon as possible; and

C. a readily discernible sign stating "SAFETY DEVICE OR INTERLOCK NOT WORKING" must be posted on the radiation source housing and at the control panel, when a safety device is bypassed.

**Statutory Authority:** *MS s 144.12*

**History:** *32 SR 777*

**Published Electronically:** *December 10, 2007*