

**4732.0700 CALIBRATIONS.**

Subpart 1. **Diagnostic radiographic system calibrations.** The registrant must ensure that corrective actions or calibrations are performed on a diagnostic radiographic system whenever that system does not meet the minimum equipment performance criteria in nationally recognized standards, such as:

- A. Code of Federal Regulations, title 21, section 1020.30;
- B. the manufacturer's specifications; or
- C. specified in part 4732.1100.

Subp. 2. **Therapeutic system calibrations.** The registrant must ensure that the corrective actions or calibrations are performed on the therapeutic equipment whenever that system does not meet the minimum equipment performance criteria in nationally recognized standards, such as:

- A. Code of Federal Regulations, title 21, section 892;
- B. the manufacturer's specifications; or
- C. specified in part 4732.1120 or 4732.1130.

Subp. 3. **Tests after change or replacement.** Calibration or an equipment performance evaluation must be performed when there is any change or replacement of components that could cause a change in the radiation output of that system.

Subp. 4. **Records.** The registrant must ensure that the records are maintained according to part 4732.0330.

**Statutory Authority:** *MS s 144.12*

**History:** *32 SR 777*

**Published Electronically:** *December 10, 2007*