

4732.0830 FLUOROSCOPIC DOSE-AREA-PRODUCT MONITOR.

All fluoroscopic equipment installed 90 days after November 5, 2007, must be equipped with a dose-area-product monitor or comparable device, capable of recording the total radiation dose received by a patient when the fluoroscopic tube is used.

Statutory Authority: *MS s 144.12*

History: *32 SR 777*

Published Electronically: *December 10, 2007*