

7017.1140 CEMS DESIGN REQUIREMENTS.

A CEMS shall be designed to complete a minimum of one cycle of sampling, analyzing, and data recording in each 15-minute period.

Statutory Authority: *MS s 116.07*

History: *23 SR 1764*

Published Electronically: *November 29, 2007*