

153.26 USE OF NITROUS OXIDE.

A podiatrist may use an oxygen-nitrous oxide mixture for sedation or analgesia only under the following conditions:

- (1) the patient is not rendered unconscious;
- (2) the nitrous oxide concentration is 70 percent of the mixture or less;
- (3) the oxygen-nitrous oxide mixture is administered only by nasal inhalator equipped with scavenging system;
- (4) an individual other than the attending podiatrist monitors the equipment and the patient during the administration of oxygen-nitrous oxide sedation and is medically directed and supervised by the podiatrist;
- (5) the podiatrist and the individual referred to in clause (4) have successfully completed, after December 31, 1986, a course in nitrous oxide sedation approved by a school of podiatric medicine;
- (6) the podiatrist and the individual referred to in clause (4) have completed, after December 31, 1986, a course in cardiopulmonary resuscitation;
- (7) oxygen-nitrous oxide is not used in conjunction with any other analgesic or sedative; and
- (8) the machine used for administering the oxygen-nitrous oxide mixture includes:
 - (i) an automatic shutoff valve in the event the oxygen concentration is less than three liters per minute;
 - (ii) an immediately available oxygen flush capability utilizing a positive pressure mask for pure oxygen delivery;
 - (iii) an automatic control over maximum nitrous oxide concentration which prevents oxygen concentration from falling below 30 percent of the mixture when nitrous oxide flow is maximal; and
 - (iv) color-coded valves, tanks, and knobs, and pin index and diameter index safety systems.

History: 1990 c 540 s 2